ENVIRONMENTAL® STANDARDS
www.envstd.com

This quality assurance (QA) review is based upon an examination of the data generated from the analyses of the samples collected as part of the:

Cumberland Water Use Survey

This review was performed with guidance from the National Functional Guidelines for Inorganic Data Review (US EPA, October 2004); the US EPA Region IV Environmental Investigations Standard Operating Procedures and Quality Assurance Manual (November 2001); and the US EPA Region IV Data Validation Standard Operating Procedures. These validation guidance documents specifically address analyses performed in accordance with the Contract Laboratory Program (CLP) analytical methods and are not completely applicable to the type of analyses and analytical protocols performed for the US EPA, SW-846, and Standard Methods utilized by the laboratory for these samples. Environmental Standards, Inc. (Environmental Standards) used professional judgment to determine the usability of the analytical results and compliance relative to the US EPA, SW-846, and Standard Methods utilized by the laboratory. This QA review was performed on the data associated with Job Number:

10638993

The findings offered in this report are based on a review of holding times and preservation, method blank results, field blank results, filter blank results, equipment blank results, tubing blank results, matrix spike/matrix spike duplicate recoveries and precision, laboratory control sample/laboratory control sample duplicate recoveries and precision, laboratory and field duplicate precision, total and dissolved results comparisons, and/or positive results between the method detection limit and quantitation limit.

The following results were qualified based on the data verification effort:

Sample	Location	Sample Type	Method	Analyte	T/D	Result	Qual	Reason Code(s)	MDL	QL	Unit

Data Qualifiers	
U*	This result should be considered “not-detected” because it was detected in a rinsate blank or laboratory blank at a similar level.
UR	Unreliable reporting limit; analyte may or may not be present in sample.
R	Unreliable positive result; analyte may or may not be present in sample.
J	Quantitation is approximate due to limitations identified during data validation.
UJ	This analyte was not detected, but the reporting limit may or may not be higher due to a bias identified during data validation.

Reason Codes and Explanations	
BE	Equipment blank contamination. The result should be considered “not-detected.”
BF	Field blank contamination. The result should be considered “not-detected.”
BL	Laboratory blank contamination. The result should be considered “not-detected.”
FD	Field duplicate imprecision.
FG	Total versus Dissolved Imprecision.
H	Holding time exceeded.
L	LCS and LCSD recoveries outside of acceptance limits, indeterminate bias.
L+	LCS and/or LCSD recoveries outside of acceptance limits. The result may be biased high.
L-	LCS and/or LCSD recoveries outside of acceptance limits. The result may be biased low.
LD	Laboratory duplicate imprecision.
LP	LCS/LCSD imprecision.
M	MS and MSD recoveries outside of acceptance limits, indeterminate bias.
M+	MS and/or MSD recoveries outside of acceptance limits. The result may be biased high.
M-	MS and/or MSD recoveries outside of acceptance limits. The result may be biased low.
MP	MS/MSD imprecision.
Q	Chemical Preservation issue.

RL	Reported Results between the MDL and RL.
T	Temperature preservation issue.
X	Percent solids < 50%.
Y+	Chemical Yield outside of acceptance limits. The result may be biased high.
Y-	Chemical yield outside of acceptance limits. The result may be biased low.

Facility: Cumberland Fossil Plant
SDG: 10638993

Lab Sample ID	10638993001
Sys Sample Code	CUF-WS-CUFPV-002-20230104
Sample Name	CUF-WS-CUFPV-002-20230104
Sample Date	1/4/2023 1:20:00 PM
Location	CUFPV-002 / CUFPV-002
Sample Type	N
Matrix	SW
Parent Sample	

Analytic Method	Chemical Name	CAS Rn	Fraction	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
EPA 200.8	Boron	7440-42-8	T	ug/L	20.3			4.5	4.5	10	Y	Yes	1	NA
	Boron, Dissolved	7440-42-8	D	ug/L	18.7			4.5	4.5	10	Y	Yes	1	NA
EPA 300.0/SW846 9056	Chloride	16887-00-6	N	mg/L	6.5			0.39	0.39	1.2	Y	Yes	1	NA
	Sulfate	14808-79-8	N	mg/L	55.3			0.43	0.43	1.2	Y	Yes	1	NA

Facility: Cumberland Fossil Plant
SDG: 10638993

Lab Sample ID	10638993002
Sys Sample Code	CUF-WS-DUP01-20230104
Sample Name	CUF-WS-DUP01-20230104
Sample Date	1/4/2023 12:00:00 AM
Location	CUFPV-002 / CUFPV-002
Sample Type	FD
Matrix	SW
Parent Sample	CUF-WS-CUFPV-002-20230104

Analytic Method	Chemical Name	CAS Rn	Fraction	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
EPA 200.8	Boron	7440-42-8	T	ug/L	18.3			4.5	4.5	10	Y	Yes	1	NA
	Boron, Dissolved	7440-42-8	D	ug/L	16.8			4.5	4.5	10	Y	Yes	1	NA
EPA 300.0/SW846 9056	Chloride	16887-00-6	N	mg/L	6.6			0.39	0.39	1.2	Y	Yes	1	NA
	Sulfate	14808-79-8	N	mg/L	55.4			0.43	0.43	1.2	Y	Yes	1	NA

Facility: Cumberland Fossil Plant
SDG: 10638993

Lab Sample ID	10638993003
Sys Sample Code	CUF-WS-FB01-20230104
Sample Name	CUF-WS-FB01-20230104
Sample Date	1/4/2023 2:40:00 PM
Location	CUFPV-002 / CUFPV-002
Sample Type	FB
Matrix	WQ
Parent Sample	

Analytic Method	Chemical Name	CAS Rn	Fraction	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
EPA 200.8	Boron	7440-42-8	T	ug/L		U		4.5	4.5	10	N	Yes	1	NA
	Boron, Dissolved	7440-42-8	D	ug/L		U		4.5	4.5	10	N	Yes	1	NA
EPA 300.0/SW846 9056	Chloride	16887-00-6	N	mg/L		U		0.39	0.39	1.2	N	Yes	1	NA
	Sulfate	14808-79-8	N	mg/L		U		0.43	0.43	1.2	N	Yes	1	NA

Lab Sample ID	10638993004
Sys Sample Code	CUF-WS-TUB01-20230104
Sample Name	CUF-WS-TUB01-20230104
Sample Date	1/4/2023 6:20:00 PM
Location	CUFPV-002 / CUFPV-002
Sample Type	TB
Matrix	WQ
Parent Sample	

Analytic Method	Chemical Name	CAS Rn	Fraction	Result Unit	Final Result	Final Qual	Reason code	Final MDL	Final RL	Final QL	Final Detect	Final Report	DF	Basis
EPA 200.8	Boron	7440-42-8	T	ug/L		U		4.5	4.5	10	N	Yes	1	NA
	Boron, Dissolved	7440-42-8	D	ug/L		U		4.5	4.5	10	N	Yes	1	NA
EPA 300.0/SW846 9056	Chloride	16887-00-6	N	mg/L		U		0.39	0.39	1.2	N	Yes	1	NA
	Sulfate	14808-79-8	N	mg/L		U		0.43	0.43	1.2	N	Yes	1	NA